

Material Specification

Declaration of Compliance with the Order EN 10204-2.1

OxyFerm Sensors

Ref 237450, 237452-237455, 237540-237544, 237713, 237715, 243100-243104, 243289

Hamilton Bonaduz AG, Bonaduz, Switzerland

hereby declares that the wetted parts fulfill the following material specifications:

| Wetted Parts | Material | Regulatory/Specification |
|--------------|---------------------------|---|
| Sensor shaft | Stainless steel 1.4435 | R _a < 0.4 μm (N5) |
| O-ring(s) | FDA: EPDM CIP: FKM | FDA 21 CFR 177.2600 EU 1935/2004 USP <88> class VI (121 °C) |

For OxyFerm Optiflow membranes see separate material specification 695179.

Hamilton Bonaduz AG



Dr. Clara Caminada
Vice President Process Analytics



Peter Pianegonda
Quality Manager Process Analytics

Bonaduz, 2017 – 11-30

REF 695174/02

HAMILTON 

Web: www.hamiltoncompany.com
USA: 800-648-5950
Europe: +41-58-610-10-10

Hamilton Americas & Pacific Rim
4970 Energy Way
Reno, Nevada 89502 USA
Tel: +1-775-858-3000
Fax: +1-775-856-7259
sales@hamiltoncompany.com

Hamilton Europe, Asia & Africa
Via Crusch 8
CH-7402 Bonaduz, Switzerland
Tel: +41-58-610-10-10
Fax: +41-58-610-00-10
contact.pa.ch@hamilton.ch